



CPC, Taiwan - QATARGAS, Qatar Correlation Test 2023



COMPARISON RESULTS BETWEEN TERMINALS

Company	Terminal	GC No.	O2	N2	CH4	C2H6	CO2	C3H8	i-C4H10	n-C4H10	i-C5H12	n-C5H12	C6H14
CPC	Yung-An	I											
		H											
	Taichung	C											
		D											
QATARGAS	South Lab	6	0.00	0.61	93.03	6.27	0.00	0.09	0.00	0.00	0.00	0.00	0.00
		7	0.00	0.61	93.03	6.27	0.00	0.09	0.00	0.00	0.00	0.00	0.00
		8	0.00	0.61	93.01	6.29	0.00	0.09	0.00	0.00	0.00	0.00	0.00
	North Lab	42	0.00	0.61	92.97	6.33	0.00	0.09	0.00	0.00	0.00	0.00	0.00
		48	0.00	0.61	92.98	6.32	0.00	0.09	0.00	0.00	0.00	0.00	0.00
		49	0.00	0.61	92.97	6.33	0.00	0.09	0.00	0.00	0.00	0.00	0.00
Max. - Min.			0.00	0.00	0.06	0.06	0.00	0.00	0.00	0.00	0.00	0.00	0.00
Tolerance			0.03	0.03	0.30	0.10	0.03	0.05	0.03	0.03	0.03	0.03	0.03